

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



Sample: BM1673 - Virus Transport Medium RT
Batch Number: 04085272
Expiry Date: 2021-06-10
Date Received: 2020-06-11
Date Tested: 2020-06-11
Sample Condition: Satisfactory
Sample Number: 380014

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax: 01324 841314
Email: info@eolabs.com

Physical	Result	Specifications	Accredited Test Method
Sterility	Conforms	Within acceptable limits	ED/SOP/005 Visual check and growth assessment following incubation for 3 days at 15-25°C and 37°C
Sterility sampling is performed in accordance with ISO 2859-1:1999. The inspection level is $\geq 0.4\%$ of the batch and the reject level ≤ 7 units depending on batch size.			
pH	7.4	7.2 +/- 0.2	ED/SOP/003 measurement by pH meter
Colour	Conforms	DE87-6 - DE90-6 Liquid / Pale Peach	ED/SOP/009 by visual observation. Range measured using Pantone® 4 colour process guide
Fill Quantity	Conforms	3ml +/- 0.3	ED/SOP/054 by gravimetric determination

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



Sample: BM1673 - Virus Transport Medium RT
Batch Number: 04085272
Expiry Date: 2021-06-10
Date Received: 2020-06-11
Date Tested: 2020-06-11
Sample Condition: Satisfactory
Sample Number: 380014

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax: 01324 841314
Email: info@eolabs.com

Accredited Test Method for Solid Media
ED/SOP/008 Quantitative evaluation using spread inoculum technique

Accredited Test Method for Liquid Media
ED/SOP/008 Semi quantitative inoculation with growth assessment from subculture

Accredited Test Method for Diluents
ED/SOP/008 Semi quantitative evaluation of viability maintenance after a holding time of 45 minutes

All of the results on this certificate of analysis relate only to the samples submitted.



Douglas Cameron
Technical Manager, E&O Diagnostics Ltd