

# Certificate of Analysis

## Client

E&O Laboratories Ltd  
Burnhouse  
Bonnybridge  
Scotland  
FK4 2HH



**Sample:** BM1673 - Virus Transport Medium RT  
**Batch Number:** 04085661  
**Expiry Date:** 2021-06-23  
**Date Received:** 2020-06-24  
**Date Tested:** 2020-06-24  
**Sample Condition:** Satisfactory  
**Sample Number:** 381062

Burnhouse, Bonnybridge  
Scotland, FK4 2HH  
Telephone: 01324 840404  
Fax:01324 841314  
Email: info@eolabs.com

Physical	Result	Specifications	Accredited Test Method
Sterility	Conforms	Within acceptable limits	ED/SOP/005 Visual check and growth assessment following incubation for 3 days at 15-25°C and 37°C
Sterility sampling is performed in accordance with ISO 2859-1:1999. The inspection level is $\geq 0.4\%$ of the batch and the reject level $\leq 7$ units depending on batch size.			
pH	7.3	7.2 +/- 0.2	ED/SOP/003 measurement by pH meter
Colour	Conforms	DE87-6 - DE90-6 Liquid / Pale Peach	ED/SOP/009 by visual observation. Range measured using Pantone® 4 colour process guide
Fill Quantity	Conforms	3ml +/- 0.3	ED/SOP/054 by gravimetric determination

# Certificate of Analysis

## Client

E&O Laboratories Ltd  
Burnhouse  
Bonnybridge  
Scotland  
FK4 2HH



**Sample:** BM1673 - Virus Transport Medium RT  
**Batch Number:** 04085661  
**Expiry Date:** 2021-06-23  
**Date Received:** 2020-06-24  
**Date Tested:** 2020-06-24  
**Sample Condition:** Satisfactory  
**Sample Number:** 381062

Burnhouse, Bonnybridge  
Scotland, FK4 2HH  
Telephone: 01324 840404  
Fax:01324 841314  
Email: info@eolabs.com

Accredited Test Method for Solid Media	Accredited Test Method for Liquid Media	Accredited Test Method for Diluents
ED/SOP/008 Quantitative evaluation using spread inoculum technique	ED/SOP/008 Semi quantitative inoculation with growth assessment from subculture	ED/SOP/008 Semi quantitative evaluation of viability maintenance after a holding time of 45 minutes

All of the results on this certificate of analysis relate only to the samples submitted.



Douglas Cameron  
Technical Manager, E&O Diagnostics Ltd