

 S GR11.6 DOC 198 REV03 OF 21/03/2022	<b>CERTIFICATE OF QUALITY ASSURANCE</b>	<b>N° 22- CRT04048</b>
--	---	--------------------------------

CODE	LOT.N°	EXP. DATE
4504C	4504C	2210652
		31/05/2025

<b>DEVICE DESCRIPTION:</b>	REGULAR SIZE TIP IN 109MM LONG DRY TUBE
----------------------------	---

**Goods have been EO treated and are guaranteed free of amplifiable human DNA, detectable DNase and RNase.**

**Goods have been manufactured according to the ISO 9001:2015 and to the ISO 18385:2016.**

**DNase test procedure**

Following internal SOP, 4N6FLOQSwabs™ are saturated with a solution containing standard DNA ladder. DNA from solution exposed to the swabs must be intact.

**RNase test procedure**

Following internal SOP, 4N6FLOQSwabs™ are saturated with a solution containing standard RNA ladder. RNA from solution exposed to the swabs must be intact.

**Human DNA contamination test procedure**

Following internal SOP swabs are extracted and a portion of the extract is tested with the Identifiler® Plus STR kit at 29 cycles on Applied Biosystems® Genetic Analyzer. A passing result is achieved if no reproducible, typeable peaks are detected above background levels within the read region for each extraction.

**All reported results were obtained at time of lot release and are guaranteed until peel opened.**

**Copan Italia SpA certifies that all product quality requirements have been met and that all information stated above is correct.**

14/07/2022

Certificate Date of Issue



Quality Representative  
Copan Italia SpA  
Via F.Perotti, 10 – 25125 Brescia (Italy)