

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



Sample: KM0006 Fastidious Anaerobe Agar
Batch Number: 00005659
Expiry Date: 16.03.2026
Date Received: 27.03.2023
Date Tested: 27.03.2023
Sample Condition: Satisfactory
Sample Number: 443794

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax:01324 841314
Email: info@eolabs.com

Product sample prepared and tested as PP1560 Fastidious Anaerobe Agar with the appropriate selective and/or enrichment supplements added.

Productivity	RGI (%)	Colonial Appearance	Colonial Appearance Specification
B.fragilis NCTC 9343	120	Grey colonies	Grey colonies
A.israelii NCTC 10236	91	Pale yellow colonies	Pale yellow colonies
P.anaerobius ATCC 27337	106	Grey colonies	Grey colonies
C.difficile NCTC 11204	107	Grey colonies	Grey colonies
C.perfringens NCTC 8237	100	Grey colonies	Grey colonies
F.nucleatum NCTC 10562	100	Pale yellow colonies	Pale yellow colonies

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



E&O Diagnostics Ltd

Sample: KM0006 Fastidious Anaerobe Agar
Batch Number: 00005659
Expiry Date: 16.03.2026
Date Received: 27.03.2023
Date Tested: 27.03.2023
Sample Condition: Satisfactory
Sample Number: 443794

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax:01324 841314
Email: info@eolabs.com

Physical	Result	Specifications
pH	7.4	7.2 ± 0.2
Appearance	Conforms	Free flowing powder
Colour of Powder	Conforms	Straw
Colour on reconstitution	Conforms	Dark red. 76-1 – 78-1

All of the results on this certificate of analysis relate only of the samples submitted.

Douglas Cameron
Technical Manager, E&O Diagnostics Ltd