

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



E&O Laboratories Ltd

Sample: KM0006 Fastidious Anaerobe Agar

Batch Number: 00012447

Expiry Date: 21/11/2027

Date Received: 09/01/2025

Date Tested: 09/01/2025

Date of Issue: 20/01/2025

Sample Condition: Satisfactory

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax:01324 841314
Email: info@eolabs.com

Product sample prepared and tested as PP1560 Fastidious Anaerobe Agar with 7% Horse Blood			
For Solid Media the RGI is a calculation of the % growth on the test media compared with the growth on a control media. The test medium must achieve an RGI between 70-120% for non-selective media / >= 50% for a selective media.			
Productivity	RGI (%)	Colonial Appearance	Colonial Appearance Specification
A.israelii NCTC 10236	93	Pale yellow colonies	Pale yellow colonies
B.fragilis NCTC 9343	96	Grey colonies	Grey colonies
C.difficile NCTC 11204	97	Grey colonies	Grey colonies
C.perfringens NCTC 8237	70	Grey colonies	Grey colonies
F.nucleatum NCTC 10562	114	Pale yellow colonies	Pale yellow colonies
P.anaerobius ATCC 27337	117	Grey colonies	Grey colonies

Physical	Result	Specification	Test Method
Colour	Conforms	Dark red. 76-1 – 78-1	ED/SOP/009 by visual observation. Range measured using Pantone guide.
pH	7.2	7.2 ± 0.2	ED/SOP/003 measurement by pH meter.

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



E&O Laboratories Ltd

Sample: KM0006 Fastidious Anaerobe Agar

Batch Number: 00012447

Expiry Date: 21/11/2027

Date Received: 09/01/2025

Date Tested: 09/01/2025

Date of Issue: 20/01/2025

Sample Condition: Satisfactory

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax: 01324 841314
Email: info@eolabs.com

All of the results on this certificate of analysis relate only to the samples submitted. Test specifications are based on ISO 11133:2014/Amd.2:2020 and internal product specifications

Douglas Cameron
Technical Manager, E&O Laboratories Ltd