

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



E&O Laboratories Ltd

Sample: KM0019 Yeast Extract Glucose Chloramphenicol Agar
Batch Number: 00014042
Expiry Date: 04/06/2027
Date Received: 01/07/2024
Date Tested: 01/07/2024
Date of Issue: 25/03/2025
Sample Condition: Satisfactory

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax:01324 841314
Email: info@eolabs.com

Product sample prepared and tested as PP1191 Yeast Extract Glucose Chloramphenicol (100mg/L) Agar			
For Solid Media the RGI is a calculation of the % growth on the test media compared with the growth on a control media. The test medium must achieve an RGI between 70-120% for non-selective media / >= 50% for a selective media.			
Productivity	RGI (%)	Colonial Appearance	Colonial Appearance Specification
A.brasiliensis NCPF 2275	102	White colonies with black pigmentation, reverse side pale yellow	White colonies with black pigmentation, reverse side pale yellow
C.albicans NCPF 3179	83	White colonies	White colonies
S.cerevisiae NCPF 3178	100	White colonies	White colonies

Test Method: ED/SOP/051 - log reduction of inoculum using drop inoculum for selective agar and fluid media. Inhibition indicates selectivity factor ≥ 2.		
Selectivity	Test Result	Specification
E.coli NCTC 12241	Inhibited	Inhibited

Physical	Result	Specification	Test Method
Colour	Conforms	Pale straw. 2-8 – 6-8	ED/SOP/009 by visual observation. Range measured using Pantone guide.
pH	6.7	6.6 ± 0.2	ED/SOP/003 measurement by pH meter.

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



E&O Laboratories Ltd

Sample: KM0019 Yeast Extract Glucose Chloramphenicol
Agar

Batch Number: 00014042

Expiry Date: 04/06/2027

Date Received: 01/07/2024

Date Tested: 01/07/2024

Date of Issue: 25/03/2025

Sample Condition: Satisfactory

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax: 01324 841314
Email: info@eolabs.com

All of the results on this certificate of analysis relate only to the samples submitted. Test specifications are based on ISO 11133:2014/Amd.2:2020 and internal product specifications

Douglas Cameron
Technical Manager, E&O Laboratories Ltd