

Liofilchem S.r.l.	CERTIFICATO CONTROLLO QUALITÀ QUALITY CONTROL CERTIFICATE	N°610505- 620505 Revisione 1 del 09.10.2013 Pag. 1 di 1
------------------------------	--	--

PRODOTTO / *PRODUCT* : **M 17 BROTH**

LOTTO / *BATCH* : **070325502**

DATA PRODUZIONE/*PRODUCTION DATE* : **03.07.2025**

DATA SCADENZA / *EXP. DATE* : **2029.06.29**

CARATTERISTICHE FISICHE / *PHYSICAL CHARACTERISTICS*

Terreno / <i>Medium</i>	Aspetto/ <i>Appearance</i>	Colore/ <i>Colour</i>
Disidratato/ <i>Dehydrated</i>	Omogeneo/ <i>Homogeneous</i>	Beige / <i>Beige</i>
Pronto / <i>Ready</i>	Limpido/ <i>limpid</i>	Ambra/ <i>Amber</i>

CARATTERISTICHE BIOLOGICHE / *BIOLOGICAL CHARACTERISTICS*

Tempo di incubazione / <i>Incubation time</i>	24 h - 48 h
Temperatura / <i>Temperature</i>	37 °C ± 2 °C
Modalità di incubazione / <i>Procedure of incubation</i>	√ O ₂ CO ₂

Medium	M 17 AGAR			
Microorganism	Theoretical results		Result of conformity	
	Growth	Colour	Growth	Colour
Lactobacillus delbrueckii ATCC 4797	+	Colorless	☒	☒
Saccharomyces cerevisiae ATCC 9763	+	White	☒	☒
pH at 25 °C	7.2 ± 0.2		7.1	
(+) = Crescita / Growth (–) = Inibizione / Inhibition (±) = Inibizione parziale/Partial inhibition				

LOTTO/BATCH : ☒ Idoneo / *Approved*

☐ Non Idoneo / *Not approved*

DATA / *DATE* : 08.07.2025

Responsabile Controllo Qualità
Quality Control Manager

(This document has been
established electronically and is valid without
signature)